

## Projects at the EPFL Center of MicroNanoTechnology

## MICROELECTRODE ARRAY BIOCHIPS FOR THE ELECTRICAL MONITORING OF 3D ENGINEERED HUMAN BRAIN TISSUES

M. Heuschkel, L. Gomez Baisac, C. Loussert-Fonta, L. Stoppini and A. Roux <u>hepia</u> Tissue Engineering Laboratory, HEPIA, HES-SO, Geneva

Project objective:

The objective of this project was the development of a screening and testing platform adapted to the electrophysiological monitoring of 3D neural tissues derived from reprogrammed human stem cells (iPSC) at air-liquid interface. The tissue is interfaced using a Micro-Electrode Array biochips consisting of a 10% porous polyimide membrane (thickness: 8  $\mu$ m) incorporating an array of 32 recording sites coated with a layer of black platinum, which is mounted on a printed circuit board allowing connection to external signal amplification and data acquisition. A fluidic channel located below the tissue allows the perfusion of nutrients as well as test compounds.





Figure 1: Realised biochip composed of a porous Figure 2: Magnified view of a recording area polyimide membrane comprising 4 recording areas with 8 recording sites at each. Figure 2: Magnified view of a recording area composed of 8 micro-electrodes coated with porous platinum black (black circles).





1

Figure 3: The membrane covers a fluidic channel<br/>allowing nutrients to reach 4 neural tissues at air-<br/>liquid interface (arrows) through its porosity.Figure 4: SEM of an electrode showing the<br/>porous structure of the platinum black<br/>coating. Neural tissue is also present at right.

Techniques employed: Sputtering: SPIDER600; Dry etching: IBE and SPTS-APS; Photolithography: Polyimide PI-2611, AZ 10XT-20 and AZ ECI 3027

Publications:

- [1] R Wertenbroek, et al., *SpikeOnChip: A Custom Embedded Platform for Neuronal Activity Recording and Analysis*, IEEE Trans Biomed Circuits Syst., **15**:4, 743-755 (2021)
- [2] C Loussert-Fonta, et al., Opening the black box of traumatic brain injury: a holistic approach combining human 3D neural tissue and an in vitro traumatic brain injury induction device, Front. Neurosci. 17:1189615 (2023)

Funding: HEPIA HES-SO (Coat&Check project, number 115179)